



March 29, 2023

VGI Medical, Inc.
% Richard Jansen
President
Silver Pine Consulting
3851 Mossy Oak Drive
Fort Myers, Florida 33905

Re: K213815

Trade/Device Name: SiJoin®T3
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: OUR
Dated: February 22, 2023
Received: February 23, 2023

Dear Richard Jansen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Colin
O'Neill -S 

Colin O'Neill, M.B.E.
Assistant Director
DHT6B: Division of Spinal Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K213815

Device Name

SiJoin®T3

Indications for Use (Describe)

The SiJoin®T3 is intended for sacroiliac joint fusion for conditions including sacroiliac joint disruptions and degenerative sacroiliitis. Two SiJoin® T3 implants must be used in the same sacroiliac joint.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”

510(k) Summary

Date Prepared: March 28, 2023
Contact: VGI Medical, LLC
11651 87th Street
Largo, FL 33773
Telephone: 727-565-1235
Fax: 727-388-0227
Regulatory Contact: Rich Jansen, Pharm. D.
Silver Pine Consulting, LLC
richj@s-pineconsulting.com

Trade Name: SiJoin[®]T3 Implant
Product Class: Class II
Classification: 21 CFR §888.3040
Common Name: Smooth or threaded metallic bone fixation fastener
Product Codes: OUR
Panel Code: 87

Indications for Use:

The SiJoin[®]T3 is intended for sacroiliac joint fusion for conditions including sacroiliac joint disruptions and degenerative sacroiliitis. Two SiJoin[®] T3 implants must be used in the same sacroiliac joint.

Device Description:

The VGI Medical, LLC SiJoin[®]T3 Implant consists of a series of titanium alloy (per ASTM F3001) implants that are designed to provide mechanical support by transfixing the sacroiliac joint while biologic fusion occurs. The device consists of the implant body and perpendicular fins to provide stability while fusion occurs. The devices are available in five sizes to accommodate varied patient's anatomy.

Predicate Device(s):

The VGI Medical, LLC SiJoin[®]T3 Implant is substantially equivalent to the Primary Predicate device, which is the Tenon Medical, Inc. Catamaran Sacroiliac Joint Fixation System (K180818).

Performance Testing Summary:

There are no ASTM or ISO standards for SI fixation currently available, so a battery of tests was designed to test clinical and biomechanical relevancy. Four biomechanical tests were completed. These were:

1. Static Vertical Shear
2. Dynamic Vertical Shear
3. Axial Pushout
4. Side-by-side cadaver range of motion study compared to a predicate device

Technological Characteristics:

VGI Medical, LLC. has compared the SiJoin®T3 Implant to the predicate devices in regards to indications for use, materials, function, sizes and mechanical test results. These comparisons demonstrate substantial equivalence to the predicate devices.

Conclusion:

VGI Medical, LLC. concludes that these SiJoin®T3 Implant devices are substantially equivalent to the predicate device and raises no new questions of safety or effectiveness.